



TEST UPDATE FIBROMETER UNITS OF MEASURE

Mary M. Mayo, PhD, DABCC, FACB, BCLD (ABB), Medical Director

Dear Valued Client,

PathGroup has been notified by our reference laboratory partner, ARUP Laboratories, that an incorrect unit of measure for Blood Urea Nitrogen (BUN) was reported using U/L instead of the correct unit of measure, mg/dL. BUN is a component of the FibroMeter assay (FIBROM) which ARUP Laboratories performs. **Results were not affected by the associated incorrect unit of measure and all scores, classifications, and overall test result interpretations remain unchanged.**

The incorrect unit of measure was corrected on October 28, 2015. All reports and information provided by ARUP Laboratories since that time reflect the proper unit of measure. We apologize for any inconvenience this may cause and assure you that test quality matters are handled as a top priority.

For further questions, please contact Client Services at 615-562-9300 or 1-888-474-5227.